
Midmark M11 - v1.0.0
Total Cycles: 3879

Name: RITTER 1
Sterilizer ID: V2178198
10/10/2023 09:40 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 155.6 F 0.0
2:00 155.1 F 1.9
4:00 164.0 F 4.8
6:00 185.8 F 8.2
8:00 208.1 F 13.0
10:00 216.9 F 6.4
12:00 231.6 F 10.8
14:00 229.8 F 7.0
16:00 243.4 F 12.6
18:00 240.7 F 10.8
20:00 253.1 F 16.9
22:00 263.9 F 23.2
24:00 271.8 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.3
1:00 272.8 F 29.3
2:00 272.3 F 29.4
3:00 271.8 F 29.1
4:00 271.8 F 28.8

Min 270.8 F 28.2
Max 273.0 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 178.6 F 0.0
5:00 226.8 F 0.0
10:00 215.5 F 0.0
15:00 211.9 F 0.0
20:00 201.2 F 0.0
25:00 181.4 F 0.0
30:00 178.1 F 0.0
Drying Complete

FILLING: 1:02
HEATING: 24:03
STERILIZING: 4:00
VENTING: 3:38
DRYING: 30:00
TOTAL CYCLE: 01:02:43
CYCLE COMPLETE

10/10/2023 10:42 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____