
Midmark M11 - v1.0.5
Total Cycles: 790

Name: RITTER2
Sterilizer ID: V2313228
02/11/2022 01:01 PM

BEGIN UNWRAPPED CYCLE
Temp: 270 Degrees F
Time: 3 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 132.0 F 0.0
2:00 130.1 F 1.0
4:00 154.4 F 3.4
6:00 164.8 F 7.6
8:00 193.9 F 13.4
10:00 217.8 F 6.8
12:00 231.8 F 12.4
14:00 235.7 F 9.9
16:00 236.1 F 8.8
18:00 249.5 F 15.5
20:00 262.2 F 22.9
22:00 272.3 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.1
1:00 272.3 F 29.3
2:00 272.9 F 29.9
3:00 271.7 F 29.2

Min 271.6 F 28.9
Max 273.0 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 226.1 F 0.0
5:00 222.5 F 0.0
10:00 210.6 F 0.0
15:00 200.3 F 0.0
20:00 201.7 F 0.0
25:00 203.6 F 0.0
30:00 209.9 F 0.0
Drying Complete

FILLING: 1:20
HEATING: 22:10
STERILIZING: 3:00
VENTING: 2:45
DRYING: 30:00
TOTAL CYCLE: 00:59:15
CYCLE COMPLETE

02/11/2022 02:00 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____