
Midmark M11 - v1.0.0
Total Cycles: 3925

Name: RITTER 1
Sterilizer ID: V2178198
10/19/2023 12:36 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 194.1 F 0.0
2:00 190.3 F 3.9
4:00 202.7 F 9.1
6:00 219.4 F 6.3
8:00 228.6 F 9.9
10:00 232.3 F 8.0
12:00 243.7 F 12.9
14:00 242.4 F 11.6
16:00 255.3 F 18.3
18:00 266.7 F 25.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.7 F 29.1
2:00 271.7 F 28.3
3:00 272.4 F 29.2
4:00 272.4 F 28.8

Min 271.4 F 28.3
Max 272.8 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.4 F 0.0
5:00 237.5 F 0.0
10:00 229.1 F 0.0
15:00 228.6 F 0.0
20:00 226.5 F 0.0
25:00 226.8 F 0.0
30:00 233.2 F 0.0

Drying Complete

FILLING: 1:49
HEATING: 19:31
STERILIZING: 4:00
VENTING: 3:27
DRYING: 30:00
TOTAL CYCLE: 00:58:48
CYCLE COMPLETE

10/19/2023 01:34 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____