

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1316

Name: RITTER2  
Sterilizer ID: V2313228  
07/19/2022 08:05 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 161.6 F 0.0  
2:00 161.7 F 0.1  
4:00 186.3 F 2.2  
6:00 199.6 F 7.3  
8:00 216.7 F 12.6  
10:00 234.0 F 9.0  
12:00 236.9 F 10.6  
14:00 239.5 F 12.3  
16:00 240.3 F 11.1  
18:00 253.4 F 17.6  
20:00 264.4 F 24.3  
22:00 271.1 F 29.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.6  
1:00 273.2 F 30.0  
2:00 271.8 F 29.9  
3:00 272.5 F 29.9  
4:00 271.8 F 29.9

Min 271.4 F 29.5  
Max 273.4 F 30.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 175.6 F 0.0  
5:00 235.1 F 0.0  
10:00 224.5 F 0.0  
15:00 233.3 F 0.0  
20:00 223.5 F 0.0  
25:00 232.2 F 0.0  
30:00 221.7 F 0.0  
Drying Complete

FILLING: 1:33  
HEATING: 22:48  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 01:01:30  
CYCLE COMPLETE

07/19/2022 09:06 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_