

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2103

Name: RITTER 1  
Sterilizer ID: V2178198  
07/08/2022 07:07 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 73.8 F 0.0  
2:00 79.3 F 0.0  
4:00 83.6 F 0.0  
6:00 107.0 F 0.7  
8:00 133.2 F 2.6  
10:00 162.7 F 5.0  
12:00 182.6 F 7.9  
14:00 201.0 F 11.2  
16:00 220.2 F 4.9  
18:00 221.7 F 8.0  
20:00 232.5 F 12.0  
22:00 232.3 F 8.2  
24:00 244.3 F 12.5  
26:00 242.2 F 11.7  
28:00 254.1 F 17.8  
30:00 264.7 F 24.0  
32:00 271.5 F 28.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.7  
1:00 272.7 F 29.5  
2:00 271.8 F 28.6  
3:00 272.3 F 29.5  
4:00 272.2 F 29.0  
  
Min 271.5 F 28.6  
Max 272.7 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 198.8 F 0.0  
5:00 237.8 F 0.0  
10:00 210.5 F 0.0  
15:00 194.6 F 0.0  
20:00 196.5 F 0.0  
25:00 189.1 F 0.0  
30:00 176.9 F 0.0  
  
Drying Complete

FILLING: 0:53  
HEATING: 32:05  
STERILIZING: 4:00  
VENTING: 3:07  
DRYING: 30:00  
TOTAL CYCLE: 01:10:04  
CYCLE COMPLETE

07/08/2022 08:17 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_