

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1468

Name: RITTER 1  
Sterilizer ID: V2178198  
12/16/2021 12:55 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 172.8 F 0.0  
2:00 182.9 F 2.6  
4:00 195.4 F 7.2  
6:00 215.1 F 12.4  
8:00 224.3 F 7.5  
10:00 237.4 F 12.9  
12:00 237.8 F 10.4  
14:00 236.0 F 8.8  
16:00 250.6 F 15.8  
18:00 262.6 F 22.7  
20:00 272.5 F 29.0

STERILIZING  
mm:ss Degrees PSI  
0:00 272.0 F 28.6  
1:00 272.1 F 29.1  
2:00 272.5 F 29.0  
3:00 271.6 F 28.5  
4:00 272.3 F 29.3  
  
Min 271.5 F 28.4  
Max 272.7 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.6 F 0.0  
5:00 193.6 F 0.0  
10:00 204.1 F 0.0  
15:00 208.9 F 0.0  
20:00 209.0 F 0.0  
25:00 217.0 F 0.0  
30:00 220.5 F 0.0  
Drying Complete

FILLING: 1:44  
HEATING: 20:08  
STERILIZING: 4:00  
VENTING: 3:20  
DRYING: 30:00  
TOTAL CYCLE: 00:59:12  
CYCLE COMPLETE

12/16/2021 01:54 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_