

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5486

Name: RITTER 1  
Sterilizer ID: V2178198  
04/08/2025 09:00 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 70.2 F 0.0  
2:00 87.2 F 0.0  
4:00 82.6 F 0.2  
6:00 115.3 F 0.4  
8:00 144.4 F 2.6  
10:00 176.0 F 5.4  
12:00 188.2 F 8.6  
14:00 206.0 F 12.2  
16:00 212.9 F 4.0  
18:00 224.8 F 7.6  
20:00 233.9 F 11.4  
22:00 230.7 F 7.3  
24:00 243.6 F 12.8  
26:00 240.8 F 10.9  
28:00 253.1 F 17.0  
30:00 263.8 F 23.2  
32:00 272.1 F 28.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.4  
1:00 272.7 F 29.0  
2:00 271.6 F 28.3  
3:00 272.7 F 29.1  
4:00 271.8 F 28.3  
  
Min 271.6 F 28.2  
Max 272.7 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 190.5 F 0.0  
5:00 196.7 F 0.0  
10:00 186.6 F 0.0  
15:00 192.2 F 0.0  
20:00 199.4 F 0.0  
25:00 204.9 F 0.0  
30:00 212.9 F 0.0  
  
Drying Complete

FILLING: 0:44  
HEATING: 32:11  
STERILIZING: 4:00  
VENTING: 3:37  
DRYING: 30:00  
TOTAL CYCLE: 01:10:33  
CYCLE COMPLETE

04/08/2025 10:10 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_