
Midmark M11 - v1.0.0
Total Cycles: 3938

Name: RITTER 1
Sterilizer ID: V2178198
10/23/2023 03:02 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 146.0 F 0.0
2:00 147.6 F 1.8
4:00 156.5 F 4.9
6:00 187.8 F 9.1
8:00 217.2 F 9.9
10:00 222.0 F 7.7
12:00 231.4 F 12.7
14:00 235.0 F 9.5
16:00 237.9 F 9.2
18:00 247.3 F 14.0
20:00 259.5 F 20.6
22:00 270.3 F 27.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.6
1:00 271.5 F 28.5
2:00 271.9 F 28.7
3:00 272.6 F 28.9
4:00 272.3 F 29.2

Min 271.2 F 28.5
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.2 F 0.0
5:00 202.4 F 0.0
10:00 198.7 F 0.0
15:00 198.9 F 0.0
20:00 197.5 F 0.0
25:00 194.6 F 0.0
30:00 192.5 F 0.0
Drying Complete

FILLING: 1:37
HEATING: 22:46
STERILIZING: 4:00
VENTING: 3:34
DRYING: 30:00
TOTAL CYCLE: 01:01:57
CYCLE COMPLETE

10/23/2023 04:03 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____