

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R750  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000539  
15:36 29/09/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:09  
PRESSURIZING 3:11  
STERILIZING 4:15  
136.4 C 224kPa 4:15  
136.4 C 225kPa 4:45  
136.4 C 223kPa 5:15  
136.1 C 222kPa 5:45  
136.3 C 224kPa 6:15  
136.4 C 221kPa 6:45  
136.1 C 222kPa 7:15  
136.4 C 225kPa 7:45  
136.4 C 225kPa 7:45  
VENTING 7:45  
AIR DRYING 8:39  
CYCLE COMPLETE 8:41

-----  
Digital Signature #  
0A313194B3E92B6F  
-----