
Midmark M11 - v1.0.5
Total Cycles: 3923

Name: RITTER2
Sterilizer ID: V2313228
09/10/2024 11:32 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 170.8 F 0.0
2:00 169.1 F 2.1
4:00 178.7 F 5.7
6:00 206.9 F 9.7
8:00 215.3 F 8.4
10:00 225.8 F 8.9
12:00 237.3 F 12.2
14:00 238.8 F 11.2
16:00 237.5 F 9.7
18:00 250.1 F 15.9
20:00 260.8 F 22.1
22:00 270.7 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.1
1:00 272.4 F 29.6
2:00 271.6 F 29.6
3:00 272.7 F 29.9
4:00 271.9 F 30.1

Min 270.9 F 29.0
Max 273.3 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.9 F 0.0
5:00 213.5 F 0.0
10:00 210.6 F 0.0
15:00 198.3 F 0.0
20:00 193.5 F 0.0
25:00 192.2 F 0.0
30:00 198.0 F 0.0
Drying Complete

FILLING: 1:36
HEATING: 22:36
STERILIZING: 4:00
VENTING: 3:41
DRYING: 30:00
TOTAL CYCLE: 01:01:53
CYCLE COMPLETE

09/10/2024 12:33 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____