
Midmark M11 - v1.0.0
Total Cycles: 1590

Name: RITTER 1
Sterilizer ID: V2178198
02/02/2022 02:28 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 158.6 F 0.0
2:00 158.4 F 2.9
4:00 171.9 F 6.8
6:00 204.0 F 11.4
8:00 215.6 F 5.5
10:00 227.6 F 10.5
12:00 231.4 F 7.1
14:00 242.6 F 12.9
16:00 241.9 F 11.2
18:00 254.6 F 17.7
20:00 265.8 F 24.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 272.8 F 29.1
2:00 271.9 F 28.4
3:00 271.9 F 28.7
4:00 272.7 F 29.0

Min 271.4 F 28.2
Max 272.9 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.8 F 0.0
5:00 226.4 F 0.0
10:00 238.9 F 0.0
15:00 232.6 F 0.0
20:00 224.9 F 0.0
25:00 225.4 F 0.0
30:00 228.5 F 0.0
Drying Complete

FILLING: 1:11
HEATING: 21:45
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 01:00:15
CYCLE COMPLETE

02/02/2022 03:28 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____