
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.9uS / 0.5ppm
CYCLE NUMBER 001460
8:49 09/30/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:46
PRESSURIZING 6:30
STERILIZING 7:59
136.4 C 220kPa 7:59
136.2 C 219kPa 8:29
135.8 C 218kPa 8:59
136.4 C 222kPa 9:29
136.1 C 219kPa 9:59
136.2 C 220kPa 10:29
136.4 C 222kPa 10:59
135.9 C 220kPa 11:29
VENTING 11:29
AIR DRYING 12:31
CYCLE COMPLETE 39:26

Digital Signature #
BDE108237602F411
