

-----  
Midmark M11 - v1.0.5  
Total Cycles: 129

Name: RITTER2  
Sterilizer ID: V2313228  
06/23/2021 08:10 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 74.9 F 0.0  
2:00 76.8 F 0.2  
4:00 91.6 F 1.1  
6:00 113.5 F 3.0  
8:00 137.4 F 5.8  
10:00 162.5 F 9.3  
12:00 185.5 F 13.3  
14:00 200.6 F 5.5  
16:00 209.4 F 9.0  
18:00 222.5 F 12.8  
20:00 234.7 F 8.9  
22:00 240.1 F 11.5  
24:00 244.1 F 12.3  
26:00 255.4 F 18.3  
28:00 264.9 F 24.5  
30:00 270.4 F 29.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 29.6  
1:00 272.2 F 29.6  
2:00 272.5 F 30.0  
3:00 272.4 F 29.6  
4:00 272.1 F 29.9

Min 271.5 F 29.5  
Max 273.0 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 195.5 F 0.0  
5:00 233.3 F 0.0  
10:00 218.3 F 0.0  
15:00 213.5 F 0.0  
20:00 191.4 F 0.0  
25:00 190.2 F 0.0  
30:00 186.0 F 0.0  
Drying Complete

FILLING: 0:59  
HEATING: 30:34  
STERILIZING: 4:00  
VENTING: 3:05  
DRYING: 30:00  
TOTAL CYCLE: 01:08:38  
CYCLE COMPLETE

06/23/2021 09:18 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_