

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2055

Name: RITTER2  
Sterilizer ID: V2313228  
02/15/2023 11:12 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 199.1 F 0.0  
2:00 198.4 F 3.4  
4:00 199.0 F 8.3  
6:00 218.5 F 10.5  
8:00 233.1 F 10.1  
10:00 233.4 F 8.2  
12:00 240.8 F 11.3  
14:00 246.3 F 13.9  
16:00 259.0 F 20.9  
18:00 270.6 F 28.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.9  
1:00 272.5 F 29.7  
2:00 271.5 F 28.9  
3:00 271.9 F 29.6  
4:00 272.3 F 29.5  
  
Min 271.0 F 28.8  
Max 272.6 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.3 F 0.0  
5:00 221.7 F 0.0  
10:00 231.2 F 0.0  
15:00 229.6 F 0.0  
20:00 229.2 F 0.0  
25:00 229.4 F 0.0  
30:00 229.3 F 0.0  
Drying Complete

FILLING: 1:46  
HEATING: 18:49  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 00:57:45  
CYCLE COMPLETE

02/15/2023 12:09 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_