
Midmark M11 - v1.0.5
Total Cycles: 974

Name: RITTER2
Sterilizer ID: V2313228
04/11/2022 11:03 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 194.0 F 0.0
2:00 189.1 F 1.1
4:00 199.5 F 6.7
6:00 217.3 F 12.2
8:00 229.0 F 8.5
10:00 238.2 F 10.9
12:00 243.9 F 13.1
14:00 244.5 F 12.8
16:00 258.0 F 20.1
18:00 269.9 F 27.7

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.9
1:00 272.1 F 29.7
2:00 272.4 F 29.2
3:00 271.4 F 29.1
4:00 272.6 F 29.6

Min 271.3 F 28.7
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.2 F 0.0
5:00 228.3 F 0.0
10:00 225.0 F 0.0
15:00 227.2 F 0.0
20:00 229.8 F 0.0
25:00 229.9 F 0.0
30:00 226.0 F 0.0
Drying Complete

FILLING: 1:44
HEATING: 18:48
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:57:43
CYCLE COMPLETE

04/11/2022 12:00 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____