

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 002102  
7:52 02/10/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:34  
PRESSURIZING 2:42  
STERILIZING 3:55  
136.0 C 220kPa 3:55  
136.2 C 223kPa 4:25  
136.4 C 224kPa 4:55  
136.4 C 225kPa 5:25  
136.4 C 224kPa 5:55  
136.2 C 223kPa 6:25  
136.4 C 223kPa 6:55  
136.3 C 222kPa 7:25  
136.3 C 222kPa 7:25  
VENTING 7:25  
AIR DRYING 8:17  
CYCLE COMPLETE 10:56  
-----

Digital Signature #  
334EFA57918B2B23  
-----