
Midmark M11 - v1.0.0
Total Cycles: 4837

Name: RITTER 1
Sterilizer ID: V2178198
08/21/2024 09:41 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 182.0 F 0.0
2:00 175.8 F 0.8
4:00 183.9 F 3.4
6:00 213.9 F 8.6
8:00 227.5 F 11.0
10:00 232.0 F 8.7
12:00 240.1 F 11.3
14:00 242.5 F 11.4
16:00 238.7 F 9.6
18:00 252.8 F 16.4
20:00 264.4 F 23.3

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.0
1:00 272.1 F 28.6
2:00 271.9 F 28.2
3:00 271.6 F 28.3
4:00 272.2 F 28.6

Min 271.1 F 27.8
Max 272.6 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.4 F 0.0
5:00 213.4 F 0.0
10:00 199.6 F 0.0
15:00 196.4 F 0.0
20:00 197.1 F 0.0
25:00 194.2 F 0.0
30:00 214.3 F 0.0
Drying Complete

FILLING: 1:05
HEATING: 21:49
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 01:00:13
CYCLE COMPLETE

08/21/2024 10:41 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____