
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 006377
10:37 28/09/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:46
PRESSURIZING 2:48
STERILIZING 3:56
136.4 C 221kPa 3:56
135.9 C 219kPa 4:26
136.4 C 224kPa 4:56
136.4 C 224kPa 5:26
136.4 C 224kPa 5:56
136.4 C 222kPa 6:26
136.4 C 222kPa 6:56
136.3 C 224kPa 7:26
136.3 C 224kPa 7:26
VENTING 7:26
AIR DRYING 8:16
CYCLE COMPLETE 8:35

Digital Signature #
032ED32FD42B2961
