

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2473

Name: RITTER2  
Sterilizer ID: V2313228  
06/06/2023 01:26 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 175.3 F 0.0  
2:00 170.9 F 3.6  
4:00 185.0 F 8.2  
6:00 210.9 F 10.5  
8:00 224.8 F 9.1  
10:00 232.9 F 9.4  
12:00 241.0 F 12.6  
14:00 241.2 F 11.6  
16:00 254.6 F 18.5  
18:00 266.4 F 25.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.3  
1:00 272.3 F 29.7  
2:00 272.1 F 29.9  
3:00 272.0 F 29.4  
4:00 272.3 F 30.1  
  
Min 271.3 F 29.2  
Max 272.7 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.7 F 0.0  
5:00 211.8 F 0.0  
10:00 224.2 F 0.0  
15:00 227.9 F 0.0  
20:00 230.4 F 0.0  
25:00 229.1 F 0.0  
30:00 227.8 F 0.0  
Drying Complete

FILLING: 1:31  
HEATING: 19:30  
STERILIZING: 4:00  
VENTING: 3:18  
DRYING: 30:00  
TOTAL CYCLE: 00:58:19  
CYCLE COMPLETE

06/06/2023 02:24 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_