
Midmark M11 - v1.0.5
Total Cycles: 521

Name: RITTER2
Sterilizer ID: V2313228
11/08/2021 10:54 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 185.4 F 0.0
2:00 180.9 F 3.0
4:00 193.1 F 7.9
6:00 211.7 F 13.5
8:00 221.9 F 9.4
10:00 235.1 F 9.3
12:00 242.8 F 13.3
14:00 243.7 F 12.7
16:00 256.2 F 19.4
18:00 267.8 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.1
1:00 272.5 F 30.3
2:00 272.1 F 30.3
3:00 271.5 F 29.8
4:00 272.2 F 29.7

Min 270.8 F 29.0
Max 273.2 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.0 F 0.0
5:00 233.6 F 0.0
10:00 206.8 F 0.0
15:00 172.4 F 0.0
20:00 166.3 F 0.0
25:00 168.4 F 0.0
30:00 176.3 F 0.0
Drying Complete

FILLING: 1:35
HEATING: 19:02
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:57:47
CYCLE COMPLETE

11/08/2021 11:58 AM
ACCEPTED: X
REJECTED:
Operator: U01