

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3535

Name: RITTER 1  
Sterilizer ID: V2178198  
07/17/2023 01:02 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 137.0 F 0.0  
2:00 141.0 F 0.6  
4:00 164.3 F 2.9  
6:00 188.7 F 6.2  
8:00 206.4 F 10.2  
10:00 220.6 F 7.8  
12:00 228.6 F 8.7  
14:00 239.3 F 13.4  
16:00 238.3 F 10.4  
18:00 235.8 F 8.3  
20:00 248.4 F 14.5  
22:00 259.3 F 20.6  
24:00 269.5 F 27.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.4  
1:00 272.7 F 29.0  
2:00 271.2 F 28.5  
3:00 272.6 F 29.0  
4:00 271.2 F 28.4

Min 271.1 F 28.3  
Max 272.8 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.1 F 0.0  
5:00 214.2 F 0.0  
10:00 235.7 F 0.0  
15:00 231.1 F 0.0  
20:00 226.7 F 0.0  
25:00 228.8 F 0.0  
30:00 226.8 F 0.0  
Drying Complete

FILLING: 0:56  
HEATING: 25:02  
STERILIZING: 4:00  
VENTING: 3:29  
DRYING: 30:00  
TOTAL CYCLE: 01:03:27  
CYCLE COMPLETE

07/17/2023 02:05 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_