
Midmark M11 - v1.0.5
Total Cycles: 3527

Name: RITTER2
Sterilizer ID: V2313228
04/15/2024 11:39 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 168.3 F 0.0
2:00 168.8 F 2.9
4:00 184.4 F 7.0
6:00 204.4 F 11.1
8:00 217.4 F 5.8
10:00 226.9 F 9.9
12:00 232.0 F 9.0
14:00 241.8 F 12.6
16:00 241.1 F 11.5
18:00 253.7 F 17.8
20:00 264.6 F 24.4

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.1
1:00 272.0 F 30.0
2:00 272.1 F 29.6
3:00 272.2 F 30.2
4:00 271.7 F 29.5

Min 270.9 F 28.9
Max 273.0 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.2 F 0.0
5:00 215.0 F 0.0
10:00 212.8 F 0.0
15:00 206.2 F 0.0
20:00 202.7 F 0.0
25:00 201.9 F 0.0
30:00 200.9 F 0.0
Drying Complete

FILLING: 1:29
HEATING: 21:54
STERILIZING: 4:00
VENTING: 3:44
DRYING: 30:00
TOTAL CYCLE: 01:01:07
CYCLE COMPLETE

04/15/2024 12:40 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____