
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 005823
10:47 23/06/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:28
PRESSURIZING 2:31
STERILIZING 3:23
136.0 C 220kPa 3:23
136.4 C 225kPa 3:53
136.2 C 223kPa 4:23
136.1 C 223kPa 4:53
136.3 C 225kPa 5:23
136.4 C 225kPa 5:53
136.2 C 222kPa 6:23
136.4 C 223kPa 6:53
136.4 C 223kPa 6:53
VENTING 6:53
AIR DRYING 7:44
CYCLE COMPLETE 7:49

Digital Signature #
71E9287BA8230F0F
