
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 007085
9:22 22/03/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:16
PRESSURIZING 2:18
STERILIZING 3:17
136.4 C 220kPa 3:17
136.2 C 223kPa 3:47
136.4 C 223kPa 4:17
136.4 C 223kPa 4:47
136.0 C 220kPa 5:17
136.4 C 220kPa 5:47
136.4 C 224kPa 6:17
136.4 C 221kPa 6:47
136.4 C 221kPa 6:47
VENTING 6:47
AIR DRYING 7:39
CYCLE COMPLETE 8:08

Digital Signature #
0751C813B542D9A6
