
Midmark M11 - v1.0.5
Total Cycles: 3317

Name: RITTER2
Sterilizer ID: V2313228
01/24/2024 02:39 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 119.1 F 0.0
2:00 124.9 F 0.9
4:00 151.6 F 3.6
6:00 182.6 F 7.4
8:00 196.0 F 11.8
10:00 215.8 F 5.8
12:00 228.5 F 10.4
14:00 230.3 F 8.0
16:00 241.7 F 12.9
18:00 242.1 F 11.8
20:00 254.4 F 18.2
22:00 265.7 F 25.0
24:00 272.1 F 29.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.4
1:00 272.7 F 29.7
2:00 272.0 F 30.0
3:00 272.1 F 29.6
4:00 272.4 F 30.1

Min 271.3 F 29.3
Max 273.3 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.7 F 0.0
5:00 219.5 F 0.0
10:00 191.8 F 0.0
15:00 190.1 F 0.0
20:00 189.0 F 0.0
25:00 188.8 F 0.0
30:00 188.4 F 0.0
Drying Complete

FILLING: 1:08
HEATING: 24:28
STERILIZING: 4:00
VENTING: 3:38
DRYING: 30:00
TOTAL CYCLE: 01:03:14
CYCLE COMPLETE

01/24/2024 03:42 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____