
Midmark M11 - v1.0.0
Total Cycles: 3678

Name: RITTER 1
Sterilizer ID: V2178198
08/18/2023 04:59 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 116.7 F 0.0
2:00 119.1 F 1.1
4:00 137.1 F 3.3
6:00 156.9 F 6.1
8:00 180.2 F 9.6
10:00 201.0 F 11.5
12:00 216.0 F 5.9
14:00 225.6 F 9.8
16:00 234.1 F 10.2
18:00 237.6 F 10.2
20:00 237.2 F 8.9
22:00 246.4 F 13.5
24:00 257.2 F 19.2
26:00 267.1 F 25.3
28:00 272.8 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.8
1:00 272.7 F 29.4
2:00 271.6 F 28.7
3:00 272.6 F 29.6
4:00 272.1 F 29.0

Min 271.4 F 28.7
Max 272.8 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 179.1 F 0.0
5:00 227.4 F 0.0
10:00 202.8 F 0.0
15:00 211.1 F 0.0
20:00 183.6 F 0.0
25:00 174.3 F 0.0
30:00 173.8 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 28:22
STERILIZING: 4:00
VENTING: 3:38
DRYING: 30:00
TOTAL CYCLE: 01:06:55
CYCLE COMPLETE

08/18/2023 06:05 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____