
Midmark M11 - v1.0.5
Total Cycles: 1555

Name: RITTER2
Sterilizer ID: V2313228
10/03/2022 02:06 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 195.7 F 0.0
2:00 192.9 F 1.1
4:00 204.0 F 6.4
6:00 226.8 F 11.9
8:00 231.2 F 8.7
10:00 239.7 F 10.5
12:00 244.2 F 13.3
14:00 244.5 F 12.8
16:00 257.7 F 20.0
18:00 269.6 F 27.5

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 29.0
1:00 272.8 F 29.8
2:00 272.1 F 29.0
3:00 272.7 F 29.8
4:00 272.5 F 29.3

Min 271.7 F 28.8
Max 273.2 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.8 F 0.0
5:00 225.7 F 0.0
10:00 247.9 F 0.0
15:00 238.0 F 0.0
20:00 223.1 F 0.0
25:00 218.3 F 0.0
30:00 217.4 F 0.0
Drying Complete

FILLING: 1:58
HEATING: 18:52
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:57:59
CYCLE COMPLETE

10/03/2022 03:03 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____