
Midmark M11 - v1.0.0
Total Cycles: 4217

Name: RITTER 1
Sterilizer ID: V2178198
01/19/2024 02:18 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 135.1 F 0.0
2:00 140.2 F 2.2
4:00 161.0 F 5.5
6:00 190.5 F 9.8
8:00 225.1 F 6.2
10:00 228.2 F 8.7
12:00 239.9 F 11.7
14:00 239.5 F 10.3
16:00 235.8 F 8.1
18:00 249.1 F 14.6
20:00 260.7 F 21.1
22:00 271.4 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.5
1:00 272.0 F 28.5
2:00 272.9 F 28.9
3:00 272.4 F 29.1
4:00 271.2 F 28.5

Min 271.2 F 28.4
Max 272.9 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 167.6 F 0.0
5:00 191.0 F 0.0
10:00 195.5 F 0.0
15:00 195.3 F 0.0
20:00 193.6 F 0.0
25:00 192.1 F 0.0
30:00 191.6 F 0.0
Drying Complete

FILLING: 1:14
HEATING: 23:21
STERILIZING: 4:00
VENTING: 3:28
DRYING: 30:00
TOTAL CYCLE: 01:02:03
CYCLE COMPLETE

01/19/2024 03:19 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____