

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2108

Name: RITTER2  
Sterilizer ID: V2313228  
03/01/2023 11:42 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 130.9 F 0.0  
2:00 138.0 F 1.9  
4:00 155.7 F 5.0  
6:00 179.5 F 9.3  
8:00 197.6 F 7.3  
10:00 217.1 F 7.6  
12:00 231.9 F 12.7  
14:00 235.1 F 9.8  
16:00 236.8 F 8.9  
18:00 247.5 F 14.6  
20:00 259.4 F 21.2  
22:00 270.1 F 28.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.2  
1:00 272.3 F 30.0  
2:00 271.8 F 29.3  
3:00 271.4 F 29.4  
4:00 272.5 F 30.0

Min 271.2 F 29.0  
Max 272.5 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 201.5 F 0.0  
5:00 205.2 F 0.0  
10:00 210.9 F 0.0  
15:00 210.7 F 0.0  
20:00 222.0 F 0.0  
25:00 223.1 F 0.0  
30:00 234.1 F 0.0  
Drying Complete

FILLING: 0:59  
HEATING: 22:57  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 30:00  
TOTAL CYCLE: 01:01:11  
CYCLE COMPLETE

03/01/2023 12:43 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_