
Midmark M11 - v1.0.0
Total Cycles: 4781

Name: RITTER 1
Sterilizer ID: V2178198
08/01/2024 09:49 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 163.7 F 0.0
2:00 165.6 F 2.5
4:00 171.0 F 6.3
6:00 189.5 F 10.0
8:00 211.9 F 7.9
10:00 222.8 F 6.1
12:00 232.5 F 10.3
14:00 237.3 F 8.3
16:00 240.1 F 11.5
18:00 237.9 F 9.5
20:00 250.4 F 15.5
22:00 261.3 F 21.8
24:00 271.6 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.7 F 29.1
2:00 271.9 F 28.4
3:00 272.0 F 28.8
4:00 272.6 F 29.0

Min 271.5 F 28.2
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.7 F 0.0
5:00 204.5 F 0.0
10:00 214.5 F 0.0
15:00 229.3 F 0.0
20:00 225.5 F 0.0
25:00 227.3 F 0.0
30:00 228.8 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 24:39
STERILIZING: 4:00
VENTING: 3:44
DRYING: 30:00
TOTAL CYCLE: 01:03:16
CYCLE COMPLETE

08/01/2024 10:52 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____