
Midmark M11 - v1.0.5
Total Cycles: 3597

Name: RITTER2
Sterilizer ID: V2313228
05/08/2024 08:38 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.8 F 0.0
2:00 81.2 F 0.0
4:00 112.4 F 0.0
6:00 134.1 F 1.4
8:00 140.3 F 3.9
10:00 169.8 F 6.8
12:00 204.5 F 10.6
14:00 216.5 F 5.1
16:00 219.9 F 6.5
18:00 230.4 F 10.3
20:00 233.8 F 9.0
22:00 239.5 F 11.3
24:00 235.8 F 9.1
26:00 249.1 F 15.3
28:00 259.9 F 21.3
30:00 269.9 F 27.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.4
1:00 272.7 F 29.5
2:00 272.7 F 30.0
3:00 271.4 F 29.4
4:00 272.9 F 29.7

Min 271.4 F 29.2
Max 273.1 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.1 F 0.0
5:00 228.3 F 0.0
10:00 234.1 F 0.0
15:00 224.5 F 0.0
20:00 216.1 F 0.0
25:00 228.9 F 0.0
30:00 201.2 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 31:44
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 01:09:55
CYCLE COMPLETE

05/08/2024 09:47 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____