
Midmark M11 - v1.0.5
Total Cycles: 4375

Name: RITTER2
Sterilizer ID: V2313228
02/24/2025 03:22 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 158.8 F 0.0
2:00 155.1 F 2.0
4:00 161.8 F 5.4
6:00 182.5 F 9.6
8:00 203.0 F 9.3
10:00 222.1 F 7.4
12:00 231.7 F 11.7
14:00 232.4 F 8.5
16:00 242.1 F 12.0
18:00 243.6 F 12.7
20:00 255.0 F 18.7
22:00 265.4 F 25.1
24:00 270.9 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 30.0
1:00 272.6 F 30.0
2:00 272.5 F 30.4
3:00 271.9 F 30.2
4:00 272.4 F 30.0

Min 271.8 F 29.8
Max 272.7 F 30.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.9 F 0.0
5:00 204.5 F 0.0
10:00 211.3 F 0.0
15:00 208.2 F 0.0
20:00 200.9 F 0.0
25:00 196.2 F 0.0
30:00 195.3 F 0.0
Drying Complete

FILLING: 1:21
HEATING: 24:43
STERILIZING: 4:00
VENTING: 3:56
DRYING: 30:00
TOTAL CYCLE: 01:04:00
CYCLE COMPLETE

02/24/2025 04:25 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____