
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 006707
11:22 07/12/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:14
PRESSURIZING 3:17
STERILIZING 4:07
136.0 C 220kPa 4:07
136.0 C 221kPa 4:37
136.3 C 222kPa 5:07
136.4 C 223kPa 5:37
136.3 C 223kPa 6:07
136.4 C 224kPa 6:37
136.4 C 218kPa 7:07
136.3 C 219kPa 7:37
136.3 C 219kPa 7:37
VENTING 7:37
AIR DRYING 8:28
CYCLE COMPLETE 8:38

Digital Signature #
7EBA8A0CD44C3AAF
