
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001423
14:15 09/09/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:23
PRESSURIZING 4:52
STERILIZING 6:06
136.0 C 220kPa 6:06
136.0 C 217kPa 6:36
136.4 C 223kPa 7:06
136.3 C 221kPa 7:36
136.3 C 219kPa 8:06
136.4 C 223kPa 8:36
136.4 C 224kPa 9:06
136.1 C 220kPa 9:36
VENTING 9:36
AIR DRYING 10:36
CYCLE COMPLETE 17:35

Digital Signature #
E7C90A35A217EC8A
