
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 001369
16:15 28/11/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:13
PRESSURIZING 4:15
STERILIZING 5:35
136.0 C 222kPa 5:35
136.3 C 222kPa 6:05
136.4 C 226kPa 6:35
136.4 C 224kPa 7:05
136.2 C 224kPa 7:35
136.4 C 226kPa 8:05
136.4 C 226kPa 8:35
136.4 C 225kPa 9:05
136.4 C 225kPa 9:05
VENTING 9:05
AIR DRYING 10:04
CYCLE COMPLETE 17:24

Digital Signature #
A799C25D2E997404
