
Midmark M11 - v1.0.0
Total Cycles: 2589

Name: RITTER 1
Sterilizer ID: V2178198
11/21/2022 04:14 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 212.2 F 0.0
2:00 215.3 F 5.1
4:00 219.7 F 11.5
6:00 229.5 F 8.1
8:00 238.8 F 11.0
10:00 243.9 F 12.8
12:00 243.5 F 12.0
14:00 257.1 F 19.1
16:00 269.0 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.3
1:00 272.6 F 29.0
2:00 271.8 F 28.4
3:00 272.0 F 28.8
4:00 272.6 F 29.0

Min 271.4 F 28.2
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.4 F 0.0
5:00 231.3 F 0.0
10:00 245.7 F 0.0
Dry Cycle Aborted

FILLING: 3:07
HEATING: 17:06
STERILIZING: 4:00
VENTING: 3:16
DRYING: 13:07
TOTAL CYCLE: 00:40:35

11/21/2022 04:54 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____