
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 001041
15:53 19/04/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:08
PRESSURIZING 3:11
STERILIZING 4:19
136.0 C 220kPa 4:19
136.2 C 223kPa 4:49
136.3 C 223kPa 5:19
136.4 C 224kPa 5:49
136.4 C 224kPa 6:19
136.3 C 224kPa 6:49
136.4 C 224kPa 7:19
136.4 C 224kPa 7:49
136.4 C 224kPa 7:49
VENTING 7:49
CYCLE ABORTED 8:16
STOP BUTTON PRESSED
NOT STERILE
END TIME 8:16

Digital Signature #
742A922149C46536
