
Midmark M11 - v1.0.5
Total Cycles: 3999

Name: RITTER2
Sterilizer ID: V2313228
10/04/2024 08:45 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.5 F 0.0
2:00 70.7 F 0.0
4:00 75.6 F 0.0
6:00 101.5 F 0.0
8:00 140.4 F 0.2
10:00 170.7 F 0.4
12:00 188.0 F 3.0
14:00 205.2 F 6.3
16:00 215.2 F 9.7
18:00 226.2 F 13.4
20:00 229.1 F 8.4
22:00 239.1 F 12.8
24:00 238.0 F 10.2
26:00 237.0 F 8.8
28:00 248.1 F 14.5
30:00 258.6 F 20.3
32:00 268.5 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.7
1:00 272.6 F 29.5
2:00 271.7 F 28.6
3:00 272.2 F 29.6
4:00 272.3 F 29.0

Min 271.4 F 28.6
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.4 F 0.0
5:00 187.4 F 0.0
10:00 188.2 F 0.0
15:00 182.9 F 0.0
20:00 181.1 F 0.0
25:00 180.3 F 0.0
30:00 182.2 F 0.0

Drying Complete

FILLING: 0:50
HEATING: 33:19
STERILIZING: 4:00
VENTING: 3:28
DRYING: 30:00
TOTAL CYCLE: 01:11:37
CYCLE COMPLETE

10/04/2024 09:56 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____