
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000317
7:20 17/08/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:42
PRESSURIZING 3:43
STERILIZING 4:49
136.4 C 220kPa 4:49
135.9 C 216kPa 5:19
136.2 C 214kPa 5:49
136.3 C 215kPa 6:19
136.2 C 216kPa 6:49
136.1 C 217kPa 7:19
136.4 C 215kPa 7:49
136.5 C 217kPa 8:19
136.5 C 217kPa 8:19
VENTING 8:19
AIR DRYING 9:11
CYCLE COMPLETE 9:43

Digital Signature #
58EDE72E95F80DFD
