
Midmark M11 - v1.0.5
Total Cycles: 3810

Name: RITTER2
Sterilizer ID: V2313228
07/24/2024 09:18 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.3 F 0.0
2:00 72.9 F 0.0
4:00 78.8 F 0.0
6:00 97.3 F 0.1
8:00 148.6 F 0.3
10:00 176.6 F 2.5
12:00 181.6 F 5.5
14:00 200.0 F 8.5
16:00 213.7 F 11.9
18:00 219.1 F 5.8
20:00 229.0 F 10.0
22:00 234.9 F 9.9
24:00 240.3 F 11.9
26:00 238.8 F 10.2
28:00 250.5 F 16.0
30:00 260.9 F 22.0
32:00 270.6 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 29.0
1:00 272.0 F 29.6
2:00 272.2 F 29.3
3:00 271.5 F 29.3
4:00 272.6 F 29.6

Min 271.3 F 28.9
Max 272.6 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.2 F 0.0
5:00 194.6 F 0.0
10:00 185.2 F 0.0
15:00 180.2 F 0.0
20:00 180.0 F 0.0
25:00 180.4 F 0.0
30:00 189.7 F 0.0

Drying Complete

FILLING: 0:51
HEATING: 32:43
STERILIZING: 4:00
VENTING: 3:39
DRYING: 30:00
TOTAL CYCLE: 01:11:12
CYCLE COMPLETE

07/24/2024 10:29 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____