
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000966
14:49 10/12/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:31
PRESSURIZING 2:38
STERILIZING 3:43
136.4 C 221kPa 3:43
136.4 C 219kPa 4:13
136.3 C 222kPa 4:43
136.4 C 223kPa 5:13
136.4 C 220kPa 5:43
136.4 C 221kPa 6:13
136.4 C 224kPa 6:43
136.0 C 222kPa 7:13
136.0 C 222kPa 7:13
VENTING 7:14
AIR DRYING 8:03
CYCLE COMPLETE 8:16

Digital Signature #
5C0DFDC18B85D219
