

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4093

Name: RITTER 1  
Sterilizer ID: V2178198  
12/06/2023 04:50 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 122.8 F 0.0  
2:00 122.1 F 1.4  
4:00 146.7 F 4.5  
6:00 169.2 F 8.0  
8:00 190.8 F 12.2  
10:00 215.7 F 4.9  
12:00 224.7 F 9.2  
14:00 236.3 F 10.7  
16:00 238.8 F 10.2  
18:00 236.5 F 8.7  
20:00 247.5 F 13.9  
22:00 258.8 F 20.0  
24:00 269.3 F 26.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.2  
1:00 272.3 F 28.5  
2:00 271.8 F 28.7  
3:00 272.8 F 28.9  
4:00 272.6 F 29.2

Min 271.0 F 28.0  
Max 273.1 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 180.4 F 0.0  
5:00 243.8 F 0.0  
10:00 223.5 F 0.0  
15:00 231.9 F 0.0  
20:00 226.6 F 0.0  
25:00 227.1 F 0.0  
30:00 226.1 F 0.0  
Drying Complete

FILLING: 1:07  
HEATING: 25:02  
STERILIZING: 4:00  
VENTING: 3:31  
DRYING: 30:00  
TOTAL CYCLE: 01:03:40  
CYCLE COMPLETE

12/06/2023 05:53 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_