
Midmark M11 - v1.0.0
Total Cycles: 1098

Name: RITTER 1
Sterilizer ID: V2178198
08/19/2021 09:57 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 184.1 F 0.0
2:00 182.8 F 3.2
4:00 199.3 F 7.8
6:00 219.9 F 9.6
8:00 232.1 F 10.8
10:00 234.7 F 8.4
12:00 239.5 F 9.8
14:00 248.7 F 14.5
16:00 261.5 F 21.7
18:00 272.6 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.1
1:00 272.0 F 28.3
2:00 272.7 F 28.8
3:00 271.5 F 28.5
4:00 272.3 F 28.4

Min 271.3 F 28.0
Max 272.8 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.0 F 0.0
5:00 234.8 F 0.0
10:00 200.8 F 0.0
15:00 184.3 F 0.0
20:00 178.2 F 0.0
25:00 177.2 F 0.0
Dry Cycle Aborted

FILLING: 0:58
HEATING: 18:15
STERILIZING: 4:00
VENTING: 3:04
DRYING: 26:42
TOTAL CYCLE: 00:52:59

08/19/2021 10:50 AM
ACCEPTED: X
REJECTED:
Operator: U02