

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3755

Name: RITTER 1  
Sterilizer ID: V2178198  
09/12/2023 07:46 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 71.4 F 0.0  
2:00 88.4 F 0.0  
4:00 101.3 F 0.0  
6:00 111.3 F 0.0  
8:00 142.0 F 0.9  
10:00 170.4 F 2.7  
12:00 187.5 F 5.2  
14:00 202.7 F 8.4  
16:00 217.3 F 11.9  
18:00 222.6 F 4.9  
20:00 232.3 F 9.0  
22:00 241.2 F 13.2  
24:00 238.7 F 10.0  
26:00 237.9 F 8.9  
28:00 247.8 F 13.9  
30:00 258.8 F 19.9  
32:00 268.9 F 26.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.0  
1:00 272.2 F 28.5  
2:00 271.5 F 28.3  
3:00 272.1 F 28.4  
4:00 271.5 F 28.3  
  
Min 271.3 F 27.9  
Max 272.5 F 28.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 177.1 F 0.0  
5:00 202.9 F 0.0  
10:00 194.9 F 0.0  
15:00 191.5 F 0.0  
20:00 203.6 F 0.0  
25:00 204.2 F 0.0  
30:00 197.4 F 0.0  
  
Drying Complete

FILLING: 0:46  
HEATING: 33:08  
STERILIZING: 4:00  
VENTING: 3:21  
DRYING: 30:00  
TOTAL CYCLE: 01:11:16  
CYCLE COMPLETE

09/12/2023 08:57 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_