

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 000054  
9:14 19/06/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:57  
PRESSURIZING 3:02  
STERILIZING 4:11  
136.4 C 223kPa 4:11  
136.4 C 224kPa 4:41  
136.4 C 224kPa 5:11  
136.2 C 222kPa 5:41  
136.2 C 223kPa 6:11  
136.4 C 221kPa 6:41  
136.2 C 223kPa 7:11  
136.3 C 223kPa 7:41  
136.3 C 223kPa 7:41  
VENTING 7:41  
AIR DRYING 8:31  
CYCLE COMPLETE 23:31  
-----

Digital Signature #  
459213E74A411F01  
-----