
Midmark M11 - v1.0.5
Total Cycles: 3888

Name: RITTER2
Sterilizer ID: V2313228
08/23/2024 10:13 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 175.7 F 0.0
2:00 172.1 F 1.4
4:00 172.4 F 4.4
6:00 197.3 F 8.7
8:00 208.6 F 12.9
10:00 219.4 F 6.6
12:00 230.2 F 10.9
14:00 231.2 F 7.7
16:00 241.6 F 12.9
18:00 241.4 F 11.5
20:00 253.0 F 17.4
22:00 263.3 F 23.6
24:00 272.8 F 29.8

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.2
1:00 272.8 F 29.8
2:00 271.6 F 29.1
3:00 271.9 F 29.9
4:00 272.3 F 29.5

Min 271.3 F 29.0
Max 272.8 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.2 F 0.0
5:00 205.5 F 0.0
10:00 204.1 F 0.0
15:00 199.6 F 0.0
20:00 199.4 F 0.0
25:00 204.2 F 0.0
30:00 208.5 F 0.0
Drying Complete

FILLING: 1:31
HEATING: 24:21
STERILIZING: 4:00
VENTING: 3:42
DRYING: 30:00
TOTAL CYCLE: 01:03:35
CYCLE COMPLETE

08/23/2024 11:16 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____