

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2868

Name: RITTER2  
Sterilizer ID: V2313228  
09/12/2023 10:18 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 168.2 F 0.0  
2:00 167.0 F 2.7  
4:00 181.4 F 6.4  
6:00 201.6 F 11.2  
8:00 221.2 F 6.2  
10:00 237.6 F 11.4  
12:00 233.6 F 8.4  
14:00 241.4 F 11.3  
16:00 245.0 F 13.1  
18:00 256.9 F 19.5  
20:00 267.9 F 26.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.9  
1:00 272.0 F 29.8  
2:00 272.1 F 29.5  
3:00 272.6 F 30.1  
4:00 271.8 F 29.5

Min 270.6 F 28.7  
Max 273.5 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 181.1 F 0.0  
5:00 236.4 F 0.0  
10:00 194.9 F 0.0  
15:00 191.6 F 0.0  
20:00 186.8 F 0.0  
25:00 184.4 F 0.0  
30:00 181.4 F 0.0  
Drying Complete

FILLING: 1:26  
HEATING: 21:14  
STERILIZING: 4:00  
VENTING: 3:37  
DRYING: 30:00  
TOTAL CYCLE: 01:00:18  
CYCLE COMPLETE

09/12/2023 11:18 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_