
Midmark M11 - v1.0.0
Total Cycles: 794

Name: RITTER 1
Sterilizer ID: V2178198
04/21/2021 12:53 PM

BEGIN PACKS CYCLE
Temp: 250 Degrees F
Time: 30 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 179.5 F 0.0
2:00 180.4 F 3.4
4:00 196.2 F 3.8
6:00 220.1 F 5.2
8:00 221.0 F 3.8
10:00 227.7 F 5.5
12:00 236.0 F 8.7
14:00 249.1 F 14.8

STERILIZING
mm:ss Degrees PSI
0:00 251.4 F 16.0
1:00 252.3 F 16.2
2:00 252.9 F 16.6
3:00 251.9 F 16.1
4:00 252.9 F 16.6
5:00 252.4 F 16.3
6:00 251.5 F 16.1
7:00 253.1 F 16.6
8:00 252.6 F 16.3
9:00 251.6 F 16.0
10:00 252.2 F 16.5
11:00 253.0 F 16.6
12:00 252.7 F 16.4
13:00 252.1 F 16.2
14:00 251.5 F 16.0
15:00 252.3 F 16.6
16:00 253.2 F 16.7
17:00 252.6 F 16.5
18:00 252.6 F 16.4
19:00 252.0 F 16.2
20:00 251.8 F 16.0
21:00 251.5 F 16.1
22:00 252.0 F 16.5
23:00 252.9 F 16.7
24:00 253.2 F 16.6
25:00 252.7 F 16.5
26:00 252.7 F 16.4
27:00 252.4 F 16.3
28:00 252.0 F 16.2
29:00 251.6 F 16.0
30:00 251.5 F 16.1

Min 251.3 F 16.0
Max 253.2 F 16.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.0 F 0.0
5:00 228.6 F 0.0
10:00 200.4 F 0.0
15:00 179.6 F 0.0
20:00 195.1 F 0.0
Dry Cycle Aborted

FILLING: 2:05
HEATING: 14:56
STERILIZING: 30:00
VENTING: 2:56
DRYING: 21:57

TOTAL CYCLE: 01:11:54

04/21/2021 02:04 PM

ACCEPTED: X

REJECTED:

Operator: U01