
Midmark M11 - v1.0.0
Total Cycles: 1259

Name: RITTER 1
Sterilizer ID: V2178198
10/13/2021 08:47 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 107.2 F 0.0
2:00 115.0 F 0.0
4:00 118.3 F 0.0
6:00 143.2 F 0.1
8:00 179.8 F 0.2
10:00 200.6 F 2.8
12:00 212.2 F 5.8
14:00 222.7 F 9.1
16:00 233.1 F 13.1
18:00 234.3 F 9.3
20:00 240.7 F 10.7
22:00 244.5 F 12.5
24:00 242.4 F 11.2
26:00 254.8 F 17.5
28:00 265.6 F 23.9

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 27.9
1:00 272.7 F 28.7
2:00 271.9 F 28.0
3:00 271.5 F 28.1
4:00 272.7 F 28.7

Min 271.2 F 27.7
Max 272.7 F 28.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.2 F 0.0
5:00 214.1 F 0.0
10:00 212.6 F 0.0
15:00 197.9 F 0.0
20:00 197.3 F 0.0
25:00 205.0 F 0.0
30:00 203.8 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 29:52
STERILIZING: 4:00
VENTING: 3:22
DRYING: 30:00
TOTAL CYCLE: 01:08:09
CYCLE COMPLETE

10/13/2021 09:55 AM
ACCEPTED: X
REJECTED:
Operator: U01