
Midmark M11 - v1.0.0
Total Cycles: 4482

Name: RITTER 1
Sterilizer ID: V2178198
04/23/2024 03:54 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 93.4 F 0.0
2:00 95.7 F 0.7
4:00 116.5 F 2.6
6:00 131.5 F 4.7
8:00 157.4 F 7.2
10:00 179.1 F 10.2
12:00 196.2 F 13.6
14:00 211.0 F 4.1
16:00 222.8 F 8.0
18:00 233.6 F 12.3
20:00 235.8 F 9.0
22:00 240.3 F 10.6
24:00 246.3 F 13.3
26:00 257.6 F 19.3
28:00 267.9 F 25.7
30:00 271.9 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.6
1:00 272.6 F 28.9
2:00 272.1 F 29.0
3:00 272.1 F 28.7
4:00 272.3 F 29.2

Min 271.5 F 28.4
Max 273.0 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.5 F 0.0
5:00 244.1 F 0.0
10:00 224.7 F 0.0
15:00 229.9 F 0.0
20:00 234.0 F 0.0
25:00 233.9 F 0.0
30:00 231.2 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 30:05
STERILIZING: 4:00
VENTING: 3:44
DRYING: 30:00
TOTAL CYCLE: 01:08:43
CYCLE COMPLETE

04/23/2024 05:02 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____